



High pressure gas supply. Typically from a gas bottle with pressure regulator.



Typically left open to Atmosphere, or connected to vacuum pump if sub-atmospheric (or near to atmosphere) test points are required.



Connect to DUT



Optional connection. Typically left open to atmosphere unless using the BGxxX module(s) in which case the REF port is tied into the Dual Volume Unit accessory, or connected to the DUT's reference port